



medical face mask



Surgical face mask with elastic band

TECHNICAL SPECIFICATIONS

According to EN 14683 this mask fulfills all criterias of a type IIR mask

- High Bacterial Filtration Efficiency (BFE) $\geq 98\%$
- Differential pressure $< 60 \text{ Pa/cm}^2$
- Splash resistance pressure $\geq 16 \text{ kPa}$
- Microbial cleanliness $\leq 30 \text{ cfu/g}$

Dimension Mask

- Dimensions: 175 x 95 mm

Packaging

- 50 pcs in a box
- Dimensions box: 185 x 100 x 80 mm
- Dimensions carton: 520 x 380 x 340 mm
- Weight carton: 7 kg (40 boxes)

CONFIGURATION OVERVIEW

This mask consists of:

- 3 layers made of
 - layer 1: polypropylene spunbond non-woven fabric
 - layer 2: polypropylene melt blown fabric
 - layer 3: polypropylene spunbond non-woven fabric
- 2 holding elastics made of polyester low elastic yarn spandex yarn
- 1 nose bridge made of polyethylen or metal strip

QUALITY ASSURANCE

- This medical face mask has been tested according to EN 14683
- NELSON Test for Bacterial Filtration Efficiency (BFE)
- NELSON Test for Differential Pressure (ΔP)
- NELSON Test for Microbial Cleanliness
- NELSON Test for Blood Penetration Resistance
- CE certification according to Directive 93/42/EEC
- Designed for medical use
- High performance

INSTRUCTIONS

1. Take the outer surface of the mask
2. While facing the outer surface of the mask away from you, loop the ear loops comfortably around your ears
3. Adjust the nose bridge on your nose, according to your face
4. Hold now the nose bridge in place, pull the bottom end of the mask under your chin to entirely cover your mouth, your nose and your chin

